

A. Aroua
H. Rickli
J.-C. Stauffer
P. Schnyder
P. R. Trueb
J.-F. Valley
P. Vock
F. R. Verdun

How to set up and apply reference levels in fluoroscopy at a national level

Received: 12 January 2006
Revised: 28 July 2006
Accepted: 25 August 2006
© Springer-Verlag 2006

A. Aroua · J.-F. Valley · F. R. Verdun
University Institute of Applied
Radiation Physics,
Lausanne, Switzerland

H. Rickli
Department of Cardiology,
Kantonsspital,
St. Gallen, Switzerland

J.-C. Stauffer
Department of Cardiology,
Vaud University Hospital,
Lausanne, Switzerland

P. Schnyder
Department of Radiology,
Vaud University Hospital,
Lausanne, Switzerland

P. R. Trueb
Federal Office of Public Health
(FOPH),
Bern, Switzerland

P. Vock
Department of Radiology,
Bern University Hospital,
Bern, Switzerland

A. Aroua (✉)
Institut Universitaire de Radiophysique
Appliquée,
Grand-Pré 1,
1007 Lausanne, Switzerland
e-mail: abbas@aroua.com
Tel.: +41-21-6233434
Fax: +41-21-6233435

Abstract A nationwide survey was launched to investigate the use of fluoroscopy and establish national reference levels (RL) for dose-intensive procedures. The 2-year investigation covered five radiology and nine cardiology departments in public hospitals and private clinics, and focused on 12 examination types: 6 diagnostic and 6 interventional. A total of 1,000 examinations was registered. Information including the fluoroscopy time (T), the number of frames (N) and the dose-area product (DAP) was provided. The data set was used to establish the distributions of T,

N and the DAP and the associated RL values. The examinations were pooled to improve the statistics. A wide variation in dose and image quality in fixed geometry was observed. As an example, the skin dose rate for abdominal examinations varied in the range of 10 to 45 mGy/min for comparable image quality. A wide variability was found for several types of examinations, mainly complex ones. DAP RLs of 210, 125, 80, 240, 440 and 110 Gy cm² were established for lower limb and iliac angiography, cerebral angiography, coronary angiography, biliary drainage and stenting, cerebral embolization and PTCA, respectively. The RL values established are compared to the data published in the literature.

Keywords Reference levels · Diagnostic and interventional radiology · Patient dose

Introduction

X-ray examinations involving fluoroscopy, particularly those performed in gastroenterology and cardiology as well as angiography and interventional procedures, con-

tribute significantly to the total collective dose due to medical exposure even if their frequency is relatively low. The last survey on the exposure of the Swiss population by X-ray imaging indicated that this contribution amounts to up to 29% [1]. Moreover, these types of examinations

deliver effective doses to the patient as high as 100 mSv and lead to extremely high entrance surface doses, up to a few Gy, resulting in a significant risk of deterministic effects.

Several international bodies seriously address the issue of radiation protection of the patient and the radiologist for dose-intensive examinations. The International Commission on Radiological Protection (ICRP) dedicated one of its publications to the means and methods to use for preventing the lesions that may be induced by interventional radiology [2]. Similarly, the European Directive 97/43 Euratom states in its article 9 that the “member States shall ensure that appropriate radiological equipment, practical techniques and ancillary equipment are used for the medical exposure (a) of children, (b) as part of a health screening programme, (c) involving high doses to the patient, such as interventional radiology, computed tomography or radiotherapy” [3].

In this context, in recent years reference levels (RLs) have become of central importance in the management of radiation doses delivered to the patient in diagnostic and interventional radiology [4, 5]. In 2000 the National Radiological Protection Board (NRPB), now part of the Health Protection Agency, reviewed the UK national reference levels for a number of examinations [6], in 2002 the Swedish Radiation Protection Authority issued regulations and general advice concerning the implementation of reference levels in Sweden [7], and in 2003 the Federal Office for Radiological Protection (BfS) introduced reference levels in Germany [8]. More and more effort is focused on dose-intensive fluoroscopy examinations. At the international level, the results of several investigations of patients’ doses have been published in the recent literature [9–28]. Since the dose-intensive examinations involving fluoroscopy may result in a significant risk of deterministic effects, notably to the skin, it is recommended that the entrance surface doses be monitored for preventing skin injuries. Some authors claim that the DAP is not a suitable quantity for such a monitoring task because of the weak correlation between the DAP and the entrance surface dose [29], and that the latter should be measured directly on the patient. But in the absence of a direct measurement, conversion factors published in the literature [19, 30–33] may be used to establish the maximum (peak) skin dose from the DAP.

Setting up RLs for dose-intensive examinations involving fluoroscopy is a difficult task due to the large variability of the fluoroscopy time and the number of images leading to a wide distribution of patient doses. This is due to several factors, including the often loose definition of the examination, differences in the techniques and protocols used, the variability in the complexity of the cases and the experience of the radiologist. In 2000 already, Marshall et al. [9] pointed out the difficulties encountered when establishing RLs for interventional radiology, such as the difficulty of defining a “typical” examination, the limited number of patients and the restriction to specialist centres. On the other hand, the large variability related to the procedures makes it mean-

ingless in the application of RLs to compare a single interventional examination to the RL value. These difficulties may lead to the perception that even the concept of RL is not appropriate for interventional radiology or at least to the view that this tool should be used with caution.

To have a reliable set of RLs for a selection of examinations, a large number of registrations must be made for each type of examination, which is time and resource consuming. As an alternative, the pooling of all the examinations registered regardless of their type is proposed to improve the accuracy of the RL values [9].

In Switzerland the strategy for establishing RLs consisted of: (1) launching nationwide dosimetric surveys in order to establish national RLs for the various radiological modalities; (2) until the results of these surveys are available, European RL values are adopted when possible [34, 35]; (3) in the absence of European data, provisional RLs were calculated from average doses established by dosimetric modelling in diagnostic and interventional radiology during the 1998 nationwide survey on the exposure of the public by radio-diagnostics in Switzerland [1, 36]. The RLs were derived from the average doses using a 1.5 multiplying factor [37].

In early 2000 the Swiss Federal Office of Public Health, in collaboration with the University Institute of Applied Radiation Physics, set up a working group on the optimisation of radiation protection in the case of dose-intensive X-ray examinations. The aim of the studies conducted by this working group is to increase the awareness of the physicians, the radiographers and the medical physicists to radiation hazards in dose-intensive examinations, in order to improve the optimisation of the radiological technique. In 2002, the working group launched a study to investigate the situation of fluoroscopy in Switzerland. The aim of this study was twofold: (1) to investigate the technique used to perform 12 common types of examinations involving fluoroscopy in hospitals and clinics performing the bulk of the dose-intensive examinations in Switzerland and (2) to collect dosimetric data in order to establish RL values for the 12 types of examinations investigated. This paper describes the method, presents the results of the investigation and comments on the implementation of the proposed RL values in the case of interventional radiology.

Methods

The investigation lasted 2 years and covered five radiology and nine cardiology departments in nine public hospitals and private clinics, and focused on 12 types of examinations (6 diagnostic and 6 interventional) after a discussion with the practitioners of the participating centres. These 12 types of examinations, which will be denoted EX01 to EX12, are presented with a short description in Table 1.

The radiology departments belong to the five Swiss university hospitals, whereas the nine cardiology depart-

Table 1 The 12 types of examinations investigated

Code	Type of examination	Description
<i>Diagnostic</i>		
EX01	Barium meal	Investigation of the upper GI tract (oesophagus and stomach)
EX02	Lower limb and iliac angiography	Investigation of the morphology of abdominal aorta, the departure of renal arteries and the arteries vascularising the lower limbs for the detection of aneurisms and stenoses of the diagnose of arterioscleroses
EX03	Cerebral angiography	Investigation of head and brain arteries (selective, four vessels)
EX04	Barium enema	Investigation of colon using a contrast medium
EX05	Coronary angiography	Examination of coronary arteries
EX06	Electrophysiology	Percutanuuous transluminal examination of cardiac potentials
<i>Interventional</i>		
EX07	Hepatic embolisation	Hepatic embolisation (e.g., using Lipiodol) through the hepatic arteries, aiming at treating a haemorrhage, a tumour or an arterio-venous shunt
EX08	Biliary drainage and stent insertion	Percutaneous transhepatic drain insertion in one of the biliary tracts
EX09	Cerebral embolisation	Therapeutic obstruction (coiling) during the treatment of arterio-venous malformation (MAV), of arterio-venous fistulas (FAV), intra-cranial aneurisms, tumours or hemorrhagic lesions
EX10	Iliac dilatation and stent insertion	Improvement of lower limbs perfusion
EX11	Percutaneous coronary intervention	Balloon-dilatation of one or several coronary segments followed or not by stenting
EX12	Cardiac thermo ablation	Local thermoablation using a catheter with a radiofrequency source of the region of the heart responsible for arrhythmias
<i>Diagnostic + interventional</i>		
EX13	Coronary angiography (EX05) + percutaneous coronary intervention (EX11)	Descriptions of EX05 and EX11 above
EX14	Electrophysiology (EX06) + cardiac thermo ablation (EX12)	Descriptions of EX06 and EX12 above

ments represent 36% of the cardiology practice in the Swiss hospitals and clinics. The selection of the departments was based on the results of the 1998 nationwide survey of the exposure of the Swiss population to medical X-rays [1], which covered university hospitals, regional hospitals and clinics as well as private practices. In the case of interventional radiology, the bulk of the examinations are performed in university hospitals (80% on average for the types of examinations considered in this study). For diagnostic radiology the fraction of the total number of examinations performed in university hospitals is lower (on average 51% for the angiographies and 23% for the barium-based examinations explored here).

Each centre was asked to register 20 examinations for each of the 12 types. Four centres provided data for two combined examinations: a procedure including coronary angiography and percutaneous coronary intervention (PCI) and another including electrophysiology and thermoablation. These two procedures will be denoted EX13 and EX14. For each examination, the centres were requested to fill in out a questionnaire containing various information (fluoroscopy time, number of analogue images and digital frames, dose-area product and difficulty of the case) on the patient undergoing this examination (age, gender, height and weight) and on the practitioner

performing the examination (medical specialty, number of years of experience and number of examinations already performed). The paper questionnaire approach proved to be time costly, and an automatic registration would be of great help in future surveys.

In addition to the data collected concerning the examinations, the 27 fluoroscopy units used in the healthcare-providing centres were investigated. First, the DAP meters of the fluoroscopy units were checked using an external reference DAP meter put on top of the measuring device of the fluoroscopy unit. Each DAP meter was checked twice: before the registration of the protocols and at the end of the study. The variations between the measurements were lower than 10%. Apart from two cases where the indication of the local DAP meter was a factor of two higher than the reference instrument, all the DAP meters gave doses within the limits set by the Swiss Ordinance on X-ray Units ($\pm 30\%$).

Unlike the American RAD-IR study [18–20] where the fluoroscopy units were similar (Siemens, with Cu filtration), a large variability in the brands and the technical specifications of the fluoroscopy units used in the participating centres was registered in our case. The units are often programmed according to the anatomical region under investigation and operate under different modes for

the same irradiation geometry. The knowledge of the parameters used for the various categories of examinations was judged necessary. The units were therefore characterised in terms of the image quality, and the patient dose associated to the various modes used. The assessment of the image quality consisted in establishing the spatial resolution limit at the middle of a Leeds TOR(CDR) Test Object [38]. The dose measurements were performed using an 11-cm³ ionisation chamber connected to a Radcal 3036 dosimeter and a 20-cm-thick PMMA phantom.

The characteristics of the fluoroscopy units (dose rate, dose per frame, image quality index and spatial resolution) were established for three imaging modes (radiography, fluoroscopy and cine), various diameters (or magnifications) of the image intensifier and various imaging frequencies and for six categories of examinations: (1) barium-based examinations, (2) bile tractus examinations, (3) abdominal and peripheral angiographies, (4) cerebral angiographies, (5) coronary angiography and percutaneous coronary intervention and (6) electro-physiology and thermo-ablation. A summary of the results of this characterisation is given in Table 2. The average values of the dose rates in the fluoroscopy mode for the various categories range from 20 mGy/min for cerebral angiographies to 42 mGy/min for coronary angiography and percutaneous coronary intervention. The dose per image in the radiography mode ranges on average from 1.6 mGy for barium-based examinations to 6 mGy for abdominal and peripheral angiographies. The category “coronary angiography and percutaneous coronary intervention” is a factor of 2 higher than the category “electro-physiology and thermo-ablation” in terms of both the dose rate (fluoroscopy) and dose per image (cine mode).

It should be mentioned that all units involved in digestive investigations (barium-based) were relatively old systems (15 to 10 years) working in a continuous fluoroscopy mode, and all units dealing with cerebral investigations (diagnostic and therapeutic) were modern

systems (5 years to less than 6 months) equipped with copper filters and working in pulse mode. For all other examinations, the units did not have a copper filter, but used a pulse mode (2 to 10 years old).

The variation of the dose rate for the same irradiation geometry was found to be of the order of a factor of 3 for the best detection at low contrast. A significant decrease of the dose rate was registered (a factor of 2.5–3 for a 20-cm PPMA slab) when adding a Cu filtration.

In order to establish reference levels, the third-quartile method is adopted in this work. It prescribes the use as a RL of a dose value corresponding to 75% of the dose distribution established by a dosimetric survey. This method requires a statistically significant set of data. In our case, the number of DAP values collected per type of examination was relatively low, ranging from 20 to 184. For this reason, an overall data analysis, pooling all the examinations registered, was preferred. The hypothesis adopted is that in relative values the distribution of the parameters of interest such as the DAP for the pooled examinations is not different from that obtained for one specific examination, which will be checked statistically using the Kolmogorov-Smirnov test. For each type of examination, the fluoroscopy time (T), the number of frames (N) and the DAP were divided by the geometric mean values related to this examinations. Then the distributions of these relative values were established in a logarithmic scale and used to determine the RL values using the 75th-percentile method. The use of the geometric mean and the log scale is justified by the fact that skewed distributions often closely fit the log-normal distribution and are related to multiplicative effects [39]. The factors influencing the parameters of interest such as the DAP (e.g., patient thickness) are of a multiplicative nature. Marshall et al. [9] have already shown that logarithmic transformation of the DAP distributions for interventional procedures produces reasonably normal distributions.

Table 2 Average values and standard deviations related to the dose rate, the image quality index and the spatial resolution associated with the 27 fluoroscopy units investigated in this study and used for the various categories of examinations involving fluoroscopy

Mode	Fluoroscopy			Radiography			Cine		
	Dose rate (mGy/min)	Image quality index	Spatial resolution (mm-1)	Dose (mGy/frame)	Image quality index	Spatial resolution (mm-1)	Dose (mGy/frame)	Image quality index	Spatial resolution (mm-1)
Barium-based examinations	26±4	8±1	0.8±0.1	1.6±0.2	12±1	1.1±0.1	–	–	–
Bile tractus examinations	37±4	10±1	1.1±0.1	3.7±0.7	17±1	1.4±0.1	–	–	–
Abdominal and peripheral angiographies	31±6	10±1	1.1±0.1	6.0±0.8	17±1	1.4±0.1	–	–	–
Cerebral angiographies	20±4	9±2	1.3±0.2	3.6±0.9	17±1	1.9±0.1	–	–	–
Coronary angiography and percutaneous coronary intervention	42±11	9±2	1.1±0.1	–	–	–	0.27±0.07	14±2	1.7±0.2
Electro-physiology and thermo-ablation	24±15	10±4	1.1±0.3	–	–	–	0.13±0.04	13±5	2.1±0.6

Results

During this study, 1,000 examinations were registered. This is a good response rate since it represents 70% of the 1,440 registrations ideally expected and corresponds to an average of 77 examinations per centre and 83 examinations per type. The age distributions of the patients for the various types of examinations showed a 14- to 28-year shift

to higher ages with regard to the age profile of the general population (average ages of 53 to 67 against an average age of 39, respectively).

The first step in order to establish RLs was to plot the distributions of the fluoroscopy time (T), the number of frames (N) and the DAP for all types of examinations considered. Figure 1 gives an illustration of such distributions for one diagnostic (coronary angiography-EX05) and




Fig. 1 Distribution of the fluoroscopy time, the number of frames and the DAP for coronary angiography (EX05) and cerebral embolisation (EX09)

one interventional (cerebral embolisation-EX09) examination. The mean values and the ranges associated with the distributions of T, N and the DAP for all the 14 types of examinations addressed in this study are shown in Table 3.

The T, N and DAP distributions for the pooled patients are shown in Fig. 2, while Fig. 3 presents the DAP distributions specific to each of the 14 types of examinations for the purpose of comparison.

From Fig. 2, the 75% values were established. They were found to be equal to the geometric mean value multiplied by a factor 1.70 for the fluoroscopy time, 1.47 for the number of frames and 1.72 for the DAP. The factors established here are used to determine the 75% values in terms of T, N and DAP for the 14 types of examinations considered. The 75% values established from the overall distribution (whole dataset) are compared to the 75% values established from the distributions specific to each type of examination (Fig. 4). The 75% values established from the overall distribution, in terms of T, N and DAP, are rounded and proposed as reference levels (RL) for the 14 types of examinations considered in this study. These RLs are shown in Table 4.

Discussion

The distributions shown in Fig. 1 (linear scale) are skewed distributions characterised by an asymmetric shape: a main peak, a tail and a few extreme values. They are typical of what is expected for examinations involving fluoroscopy.

Table 3 reveals interesting features about the use of fluoroscopy in Switzerland. For barium meal (EX01) and barium enema (EX04) examinations, fluoroscopy times higher than 20 min are registered in some cases. Moreover, more than 100 frames are acquired in some barium meal procedures. This leads to DAPs as high as 400 Gy cm². This

seems too high for this kind of diagnostic examination when compared to the data published in the literature, notably the UK data [6], with an average DAP of 10.3 and 23.5 vs. 67 and 114 Gy cm² in this work for barium meal and barium enema, respectively. However, the high values registered for this examination may be explained by several factors. Firstly, the definition of barium meal is quite loose and may refer to examinations covering very different areas of the digestive tract. Moreover, this type of examination is often performed on old units without Cu filtration. It is often carried out by junior radiologists with little training who lack experience on how to use the unit properly and to optimise parameters such as the frame rate and the collimation.

Concerning lower limb angiographies (EX02), although the average fluoroscopy time is about 7 min, some extreme values higher than 20 min of fluoroscopy (up to 46 min) are registered, which may be related to difficult cases. DAPs as high as several hundred Gy cm² with an average of about 180 are involved in this type of examination, for which RL values of 33 Gy cm² [6] and 85 Gy cm² [8] are reported in the UK and Germany, respectively. This examination includes iliac angiography and can involve a wide range of vessels. It may or may not follow a CT or an MRI procedure and thus corresponds to a wide range of selectivity of the examination. The definition given in the questionnaire for this type of examination may appear not specific enough. It should be noted that this type of examination may be replaced in the future by CT angiography [40, 41].

Similarly, cerebral angiographies (EX03) are characterised by an average fluoroscopy time of 13 min, but much higher values have been found. The associated DAP can be up to 1,000 Gy cm² with an average of 121 Gy cm². For this type of examination, Brambilla et al. [17] reported a range of DAPs from 19 to 401 Gy cm², with an average of 158 Gy cm². A great variability is found for this type of procedure performed mainly at university hospitals.

Table 3 Mean values and ranges of the fluoroscopy time, number of frames and DAP for the various types of examinations

Examination code	Number of examinations	Fluoroscopy time (min), mean (range)	Number of frames, mean (range)	DAP (Gy cm ²), mean (range)
EX01	89	5.9 (0.8–35)	44 (0–201)	67 (3–441)
EX02	94	6.6 (0.2–46.7)	126 (2–372)	178 (8–747)
EX03	91	12.6 (2–85)	679 (32–5486)	121 (8–996)
EX04	41	7.8 (1.5–23.6)	17 (0–34)	114 (20–564)
EX05	184	5.7 (0.6–28.1)	1060 (10–4490)	57 (7–240)
EX06	24	3.7 (0.1–36.7)	38 (0–670)	13 (1–123)
EX07	70	22 (3.2–66)	140 (9–612)	463 (54–1703)
EX08	56	21.1 (1.3–82.5)	31 (0–119)	244 (5–1375)
EX09	58	36.5 (3.3–134)	760 (60–3348)	335 (24–1345)
EX10	72	19.2 (2–66)	158 (40–380)	344 (36–1122)
EX11	80	15.1 (2.5–49)	1,238 (259–7969)	82 (3–283)
EX12	42	22.1 (0.8–74.6)	162 (0–1270)	128 (7–495)
EX13	78	16.5 (2.6–86)	2,199 (360–6833)	178 (27–487)
EX14	21	32.4 (2.4–113)	0 (0–0)	348 (8–1442)




Fig. 2 Distribution of the fluoroscopy time, the number of frames and the DAP in relative values (the absolute value is divided by the geometric mean for the given type examination) in linear (left) and


logarithmic (right) scales. These distributions are associated with 995 examinations (14 types)

Although a specific protocol was sent to the participants, some of them have obviously provided data concerning selective cerebral angiographies. The fluoroscopy times may seem quite long (85 min fluoroscopy in one case), but cerebral angiography is a complex procedure with significant implications on the quality of life of the patient who usually is in a critical state. This examination is usually performed by senior radiologists and hence the high doses are due less to a lack of training than to the complexity of the case. In the future, the use of CT and MRI for this kind

of investigation will reduce the collective dose associated with this type of fluoroscopy examination.

Coronary angiography (EX05) and percutaneous coronary intervention (EX11) and the procedure combining the two (EX13) are characterised by a great dispersion of the number of frames with high average values of 1,000–2,000. The associated DAPs range from 57 to 178 Gy cm². This is comparable with the average values published in the literature and compiled by Maeder et al. [42]: 6.2–109 Gy cm² for coronary angiography and 6.7–163 Gy cm² for percutaneous coronary intervention.


Fig. 3 Distributions of the DAP in relative values (the absolute value is divided by the geometric mean for the given type examination) in logarithmic scale for each of the 14 types of examinations



Regarding electrophysiological study (EX06), thermoablation (EX12) and the procedure combining the two (EX14), they are characterised by almost no frame acquisition. Two of the three centres that provided data

for these types of examinations used fluoroscopy exclusively. Thermo-ablation involves a fluoroscopy time six times higher than electrophysiological study on average. The DAPs involved are quite high in some cases (up to

Fig. 4 Ratio of the 75% values of the fluoroscopy time, the number of frames and the DAP for the various types of examinations, established in two ways: from the distribution related to the whole set of data (overall) and from the distribution related to the type of examination (specific). The number of centres involved and the number of examinations registered are given between brackets on the x axis



about 500 Gy cm²), with average values of 13 and 128 Gy cm² for electrophysiology and thermo-ablation, respectively. In the UK average values of 17.1 and 33.1 Gy cm² are reported by NRPB [6] for these types of examinations, respectively, whereas McFadden et al. reported an average DAP of 123 Gy cm² and a maximum value of 430 Gy cm² for RF ablation [28].

Hepatic embolisation (EX07) and biliary drainage (EX08) use on average the same fluoroscopy time, but on average the former requires 4.5 times more frames than the latter. The DAPs registered are on average 463 and 244 Gy cm², respectively. These types of examinations are

performed on units without Cu filtration. Miller et al. [18] reported average DAPs of 282 and 70.6 Gy cm² for hepatic embolisation and biliary drainage, respectively, which is about a factor of 2 lower than our values. It is worth mentioning that for hepatic embolisation the average T and N in this study are 22 min and 140 frames, whereas Miller et al. reported 16.8 min and 216 frames, respectively. The lower value for the DAP is due to the fact that in the RAD-IR study the fluoroscopy units were used with Cu filtration.

It appears from Table 3 that a fluoroscopy time about three times higher is required for cerebral embolisation (EX09-interventional) than for cerebral angiography (EX03-diag-

Table 4 Reference levels (RL) for the 14 types of examinations considered

Examination code	Type of examination	T(min)	Number of frames	DAP (Gy cm ²)
<i>Diagnostic</i>				
EX01	Barium meal	7	40	60
EX02	Lower limb and iliac angiography	8	150	210
EX03	Cerebral angiography	15	480	125
EX04	Barium enema	10	25	150
EX05	Coronary angiography	7	1,400	80
EX06	Electrophysiology	1	480	10
<i>Interventional</i>				
EX07	Hepatic embolisation	30	160	620
EX08	Biliary drainage and stent insertion	25	30	240
EX09	Cerebral embolisation	50	800	440
EX10	Iliac dilatation and stent insertion	25	200	460
EX11	Percutaneous coronary intervention	20	1,500	110
EX12	Cardiac thermo-ablation	20	800	140
<i>Combined</i>				
EX13	Coronary angiography + percutaneous coronary intervention	20	2,800	260
EX14	Electrophysiology + cardiac thermo-ablation	30	-	280

nostic). The average DAP related to cerebral embolisation is 335 Gy cm², but higher values up to 1,345 Gy cm² are registered. Cerebral embolisations are performed in three specialised centres on modern units with Cu filtration; they still deliver high doses due to the complexity of the procedure. The data published in the literature are comparable to the figures established in the work. Brambilla et al. [17] reported an average DAP of 329 Gy cm² and a maximum of 1,221 Gy cm² for this type of examination. On the other hand, Miller et al. [18] reported an average value of 320 Gy cm² and a maximum of 1,351 Gy cm².

Similarly, the mean fluoroscopy times for iliac dilatation and stent insertion (EX10-interventional) differ by a factor of three from lower limb angiography (EX2-diagnostic). For iliac dilatation and stent insertion, a large variability of the field area is registered, depending on the localisation of the lesion. The DAP related to iliac dilatation and stent insertion is 344 Gy cm² on average, with a maximum of 1,122 Gy

cm². Miller et al. [18] reported an average DAP of 203 Gy cm² and a maximum of 886 Gy cm². The DAPs registered in the RAD-IR are therefore 70% lower on average than the data established in this work; also the fluoroscopy times are comparable and the number of frames is lower in our case: on average 19.2 min and 158 frames in this investigation vs. 17.4 min and 227 frames in the RAD-IR survey. This is due to the use of Cu filtration in the latter case.

Figure 2 shows that the distributions of the fluoroscopy time, the number of frames and the DAP are slightly different, but with similar shape. The full width at half maximum related to the distribution on a logarithmic scale gives a good indication of the data variability. The shape of the distribution of the pooled examinations (Fig. 2) is similar to the distributions related to specific examinations (Fig. 3), and this holds for the T and N distributions, too. This similarity was confirmed by the Kolmogorov-Smirnov test (a P-value between 12.4% and 62% for ten spe-

Table 5 RL comparisons with values published in the literature

Type of examination	Fluoroscopy time (min)	Number of frames	DAP (Gy cm ²)	Reference
EX01: barium meal	7	40	60	This work
	–	–	70	Veit and Bauer [8]
	2.3	–	13	Hart et al. [6]
EX02: lower limb and iliac angiography	8	150	210	This work
	–	–	85	Veit and Bauer [8]
	5	–	33	Hart et al. [6]
EX03: cerebral angiography	15	480	125	This work
	17.5 ^a	–	198	Brambilla et al. [17]
EX04: barium enema	10	25	150	This work
	–	–	70	Veit and Bauer [8]
	–	–	50	Holm and Leitz [7]
	2.7	–	31	Hart et al. [6]
EX05: coronary angiography	7	1,400	80	This work
	–	–	80	Holm and Leitz [7]
	6	1,270	57	Neofotistou et al. [11]
EX07: hepatic embolisation	30	160	620	This work
	21.5 ^a	266 ^a	353 ^a	Miller et al. [18]
EX08: biliary drainage	25	30	240	This work
	17	–	54	Hart et al. [6]
	30 ^a	18.4 ^a	88 ^a	Miller et al. [18]
	17.3 ^a	18 ^a	38.3 ^a	Stratakis et al. [26]
EX09: cerebral embolisation	50	800	440	This work
	12.7 ^a	–	338	Brambilla et al. [17]
	112 ^a	1,309 ^a	400 ^a	Miller et al. [18]
EX11: percutaneous coronary intervention	20	1,500	110	This work
	16	1,355	94	Neofotistou et al. [11]
	23.1	1,536	69	Prieto et al. [43]
EX12: cardiac thermo-ablation	20	800	140	This work
	86 ^a	–	154 ^{a,b}	McFadden et al. [28]

^aEstablished from the mean value

^bThe authors proposed three RLs of 100, 300 and 550 Gy cm² to provide an indication of the risk of the patient suffering transient erythema, main erythema and moist desquamation, respectively

cific distributions, between 1.7% and 4.2% for three specific distributions and 0.77% for one specific distribution).

The set of RLs proposed in this study are compared to data published in the literature for ten types of examinations as shown in Table 5. In some cases the RL was established by multiplying the published mean value by a factor as described above. Table 5 reveals significant differences with the UK data, which are noticeably low. The RLs established in this work are higher than the values proposed in Germany for barium meal and enema and for lower limb angiography. But our data compare well with the RLs published by Brambilla et al. [17], Neofotistou et al. [11], Miller et al. [18], McFadden et al. [28] and Prieto et al. [43].

Conclusion

This study allowed the collection of useful data concerning 1,000 examinations corresponding to 14 types: six diagnostic, six interventional and two combined. A large variability of the dose was registered related to several factors such as the degree of complexity of the examination, the experience of the practitioner and differences in the definition of the examination and in the technique used.

The results of this investigation indicate that interventional procedures deliver on average doses three times higher than diagnostic examinations. The dosimetric data base set up in this work allowed the establishment of national RL values for the 14 types of examinations investigated. The data collected (T, N and DAP) were divided by the mean value for the given type of examination. The distribution of the whole set of relative values was then used to deduce the RL using the third-quartile method. The RLs for T range from 1 min for electrophysiology to 50 min for cerebral embolisation, thus covering one order of magnitude. The RLs for N range from 25

frames for barium enema to 2,800 frames for the combined procedure coronary angiography + percutaneous coronary intervention, i.e., more than two orders of magnitude. Concerning the DAP, the RLs range from 10 Gy cm² for electrophysiology to 620 Gy cm² for hepatic embolisation, i.e., a difference of a factor of 40.

The RL concept cannot be applied to a specific examination of one patient for the types of examinations considered here, but it is highly recommended to perform a check, for instance, after every 20 examinations, to see if on average the RL is exceeded systematically or not, and to take the appropriate measures if necessary. Moreover, in order to prevent skin injuries, the entrance surface doses should be monitored either by direct measurements or by converting the DAP using appropriate conversion factors published in the literature, and patients who have undergone examinations with an entrance skin dose above 1 Gy, particularly children, should be followed up.

The 75th-percentile values established in this work were found to be higher than the RLs recommended in Germany and the UK. This might be due to the fact that the survey covered university hospitals mainly, where the proportion of “heavy” cases might be relatively high and where there is a high number of junior radiologists undergoing their training. These high dose values should prompt a discussion between the participating radiologists concerning the specifications of the units, the choice of the protocols and technical parameters and the training of the staff. In particular, the use of Cu filtration should be discussed within the community of radiologists.

Acknowledgments This research project was funded by the Swiss Federal Office of Public Health (contract no. 99.000.729). The authors are grateful to all the radiologists, cardiologists, medical physicists and radiographers who collaborated with us or provided assistance in the participating hospitals and clinics during this survey.

References

1. Aroua A, Burnand B, Decka I, Vader JP, Valley JF (2002) Nation-wide survey on radiation doses in diagnostic and interventional radiology in Switzerland in 1998. *Health Phys* 83(1):46–55
2. International Commission on Radiological Protection (2000) Avoidance of radiation injuries from medical interventional procedures. ICRP Publication 85
3. European Union (1997) On health protection of individuals against the dangers of ionising radiation in relation to medical exposure, Council Directive 97/43/Euratom, Official J Eur Commun L 180:22–27
4. International Commission on Radiological Protection (2001) Diagnostic reference levels in medical imaging. ICRP Committee 3, (draft)
5. European Commission (1999) Guidance on diagnostic reference levels (DRLs) for medical exposures. *Radiat Protect* 109
6. Hart D, Hiller MC, Wall BF (2002) Doses to patients from medical X-ray examinations in the UK-2000 Review. Report NRPB-W14. National Radiological Protection Board, Didcot
7. Holm LE, Leitz W (2002) The Swedish Radiation Protection Authority's regulations and general advice on diagnostic standard doses and reference levels within medical X-ray diagnostics. SSI FS 2002:2 Statens Strålskyddsinstitut Stockholm

8. Veit R, Bauer B (2003) Introduction of diagnostic reference levels into diagnostic radiology in Germany. Internal report. Federal Office for Radiological Protection, Salzgitter
9. Marshall NW, Chapple CL, Kotre CJ (2000) Diagnostic reference levels in interventional radiology. *Phys Med Biol* 45:3833–3846
10. Van de Putte S, Verhaegen F, Taeymans Y, Thierens H (2000) Correlation of patient skin doses in cardiac interventional radiology with dose-area product. *Br J Radiol* 73:504–513
11. Neofotistou V, Vano E, Padovani R, Kotre J, Dowling A, Toivonen M, Kottou S, Tsapaki V, Willis S, Bernardi G, Faulkner K (2003) Preliminary reference levels in interventional cardiology. *Eur Radiol* 13:2259–2263
12. Martin CJ (2004) A review of factors affecting patient doses for barium enemas and meals. *Br J Radiol* 77:864–868
13. Ben Salem D, Martin D, Aho LS, Walker PM, Lalonde A, Brunotte F, Krause D, Ricolfi F (2004) Analysis of variation in delivered dose in diagnostic and therapeutic cerebral angiography. (in French). *J Neuroradiol* 31:379–383
14. Bor D, Sancak T, Olgar T, Elcim Y, Adanali A, Sanlidilek U, Akyar S (2004) Comparison of effective doses obtained from dose-area product and air kerma measurements in interventional radiology. *Br J Radiol* 77:315–322
15. Delichas MG, Psarrakos K, Giannoglou G, Molyvda-Athanasopoulou E, Hatzioannou K, Papanastassiou E (2005) Skin doses to patients undergoing coronary angiography in a Greek hospital. *Radiat Prot Dosim* 113 (4):449–452
16. Persliden J (2005) Patient and staff doses in interventional X-ray procedures in Sweden. *Radiat Prot Dosim* 114(1–3):150–157
17. Brambilla M, Marano G, Dominietto M, Cotroneo AR, Carriero A (2004) Patient radiation doses and reference levels in interventional radiology. *Radiol Med* 107:408–418
18. Miller DL, Balter S, Cole PE, Lu HT, Schuler BA, Geisinger M, Berenstein A, Albert R, Georgia JD, Noonan PT, Cardella JF, St. George J, Russell EJ, Malisch TW, Vogelzang RL, Miller GL III, Anderson J (2003) Radiation doses in interventional radiology procedures: The RAD-IR study. Part I: overall measures of dose. *J Vasc Interv Radiol* 14:711–727
19. Miller DL, Balter S, Cole PE, Lu HT, Berenstein A, Albert R, Schuler BA, Georgia JD, Noonan PT, Russell EJ, Malisch TW, Vogelzang RL, Geisinger M, Cardella JF, St. George J, Miller GL III, Anderson J (2003) Radiation doses in interventional radiology procedures: The RAD-IR study. Part II: skin dose. *J Vasc Interv Radiol* 14:977–990
20. Balter S, Schuler BA, Miller DL, Cole PE, Lu HT, Berenstein A, Albert R, Georgia JD, Noonan PT, Russell EJ, Malisch TW, Vogelzang RL, Geisinger M, Cardella JF, St. George J, Miller GL III, Anderson J (2004) Radiation doses in interventional radiology procedures: The RAD-IR study. Part III: dosimetric performance of the interventional fluoroscopy units. *J Vasc Interv Radiol* 15:919–926
21. Davies AG, Cowen AR, Kengyelics SM, Moore J, Pepper C, Cowan C, Sivanathan MU (2006) X-ray dose reduction in fluoroscopically guided electrophysiology procedures. *Pacing Clin Electrophysiol* 29(3):262–271
22. Perisinakis K, Raissaki M, Damilakis J, Stratakis J, Neratzoulakis J, Gourtsoyiannis N (2006) Fluoroscopy-controlled voiding cystourethrography in infants and children: are the radiation risks trivial? *Eur Radiol* 16 (4):846–851
23. Peterzol A, Quai E, Padovani R, Bernardi G, Kotre CJ, Dowling A (2005) Reference levels in PTCA as a function of procedure complexity. *Radiat Prot Dosim* 117(1–3):54–58
24. Dragusin O, Desmet W, Heidbuchel H, Padovani R, Bosmans H (2005) Radiation dose levels during interventional cardiology procedures in a tertiary care hospital. *Radiat Prot Dosim* 117(1–3):231–235
25. Goni H, Tsalafoutas IA, Tzortzis G, Pappas P, Bouzas N, Loulakas J, Georgiou A, Georgiou E, Yakoumakis EN (2005) Radiation doses to patients from digital subtraction angiography. *Radiat Prot Dosim* 117(1–3):251–255
26. Geijer H, Persliden J (2004) Radiation exposure and patient experience during percutaneous coronary intervention using radial and femoral artery access. *Eur Radiol* 14(9):1674–1680
27. Stratakis J, Damilakis J, Hatzidakis A, Perisinakis K, Gourtsoyiannis N (2006) Radiation dose and risk from fluoroscopically guided percutaneous transhepatic biliary procedures. *J Vasc Interv Radiol* 17(1):77–84
28. McFadden SL, Mooney RB, Shepherd PH (2002) X-ray dose and associated risks from radiofrequency catheter ablation procedures. *Br J Radiol* 75:253–265
29. Chu RY, Thomas G, Maqbool F (2006) Skin entrance radiation dose in an interventional radiology procedure. *Health Physics* 91(1):41–46
30. Hansson B, Karambatsakidou A (2000) Relationships between entrance skin dose, effective dose and dose area product for patients in diagnostic and interventional procedures. *Radiat Prot Dosim* 90(1–2):141–144
31. Karambatsakidou A, Tornvall P, Saleh N, Chouliaras T, Löfberg P-O, Fransson A (2005) Skin dose alarm levels in cardiac angiography procedures: is a single DAP value sufficient? *Br J Radiol* 78:803–809
32. Chida K, Saito H, Otani H, Kohzuki M, Takahashi S, Yamada S, Shirato K, Zuguchi M (2006) Relationship between fluoroscopic time, dose-area product, body weight, and maximum radiation skin dose in cardiac interventional procedures. *Am J Roentgenol* 186(3):774–778

-
33. Vano E, Gonzalez L, Ten JE, Fernandez JM, Guibelalde E, Macaya C (2001) Skin dose and dose-area product values for interventional cardiology procedures. *Br J Radiol* 74:48–55
 34. Office fédéral de la santé publique. Division de radioprotection. Niveaux de référence diagnostiques (NRD) pour les examens de radiographie. Directive R-08-04. Etablie le 7 avril 2004. Accessible en ligne sur: <http://www.bag.admin.ch/strahlen/lois/pdf/R-08-04-mf.pdf>
 35. Accessible on: <http://www.bag.admin.ch/strahlen/lois/pdf/NRDCalc1.2.xls>
 36. Aroua A, Decka I, Burnand B, Vader JP, Valley JF (2002) Dosimetric aspects of a national survey of diagnostic and interventional radiology in Switzerland. *Med Phys* 29(10):2247–2259
 37. Aroua A, Besançon A, Buchillier-Decka I, Trueb P, Valley JF, Verdun FR, Zeller W (2004) Adult reference levels in diagnostic and interventional radiology for temporary use in Switzerland. *Radiat Prot Dosim* 111(3):289–295
 38. Llorca AL, Guibelalde E, Vañó N, Ruiz MJ (1993) Analysis of image quality parameters using a combination of an ANSI type phantom and the Leeds TOR(CDR) Test object in simulations of simple examinations. *Radiat Prot Dosim* 49:47–49
 39. Limpert E, Stahel WA, Abbt M (2001) Log-normal distributions across the sciences: keys and clues. *Bioscience* 51(5):341–352
 40. Kock M, Adriaansen M, Pattynama P, Van Sambeek M, Van Urk H, Stijnen T, Hunink M (2005) DSA versus multi-detector row CT angiography in peripheral arterial disease: randomized controlled trial. *Radiology* 237:727–737
 41. Fraioli F, Catalano C, Napoli A, Francone M, Venditti F, Danti M, Pediconi F, Passariello R (2006) Low-dose multidetector-row CT angiography of the infra-renal aorta and lower extremity vessels: image quality and diagnostic accuracy in comparison with standard DSA. *Eur Radiol* 16(1):137–146
 42. Maeder M, Verdun FR, Stauffer JC, Ammann P, Rickli H (2005) Radiation exposure and radiation protection in interventional cardiology. *Kardiovaskuläre Medizin* 8:124–132
 43. Prieto C, Vano E, Fernandez JM, Galvan C, Sabate M, Gonzalez L, Martinez D (2006) Six years experience in intracoronary brachytherapy procedures: patient doses from fluoroscopy. *Br J Radiol* DOI 10.1259/bjr/75766147. Published online before print June 22